



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

1184

DATE: Mar 13, 2006 INITIALS: DEC OWNER: WSP
DISTRICT: 8 ADDRESS: _____
LOCATION: 7th Shop COUNTY: _____
MAKE: Haenai TYPE: port wheel weighing ACCURACY CLASS: TTT
MODEL: 101 ID. NO: 15836 CAPACITY: 20,000 lb SECTIONS: 1
MINIMUM GRADUATIONS: 50 lb REG. ELEMENT: Nil PLATFORM: 15" X 26"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Center</u>	<u>5000</u>	<u>5000</u>	<u>0</u>				
<u>"</u>	<u>10000</u>	<u>10050</u>	<u>+50</u>				
<u>"</u>	<u>15000</u>	<u>15050</u>	<u>+50</u>				
<u>Center</u>	<u>20000</u>	<u>20000</u>	<u>0</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 195-77


I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Crona Date: 3/28/06

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before
me on March 28, 2006 by

Michael J. Evans
(Signature of notary public)

Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS